

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV060920\
 Data File : VV016721.D
 Acq On : 09 Jun 2020 16:49
 Operator : SY/MD
 Sample : VV0609WBL02
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK76

Quant Time: Jun 10 03:51:32 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jun 10 03:42:58 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	362631	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	341344	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	167763	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	91350	35.70	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	71.40%
7) Chloroethane-d5	1.57	69	74217	48.34	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	96.68%
11) 1,1-Dichloroethene-d2	2.12	63	135941	31.64	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	63.28%
21) 2-Butanone-d5	3.89	46	185860	98.14	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	98.14%
24) Chloroform-d	4.38	84	228124	43.84	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.68%
26) 1,2-Dichloroethane-d4	5.05	65	163215	46.12	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.24%
32) Benzene-d6	5.07	84	442361	44.96	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.92%
36) 1,2-Dichloropropane-d6	6.09	67	142613	46.50	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	93.00%
41) Toluene-d8	7.34	98	396710	43.94	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	87.88%
43) trans-1,3-Dichloropropene-	7.64	79	66208	41.45	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	82.90%
47) 2-Hexanone-d5	8.11	63	124937	93.43	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	93.43%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	191873	47.26	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	94.52%
64) 1,2-Dichlorobenzene-d4	11.65	152	169057	47.22	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.44%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV060920\
Data File : VV016721.D
Acq On : 09 Jun 2020 16:49
Operator : SY/MD
Sample : VV0609WBL02
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK76

Quant Time: Jun 10 03:51:32 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jun 10 03:42:58 2020
Response via : Initial Calibration

