

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV060920\
 Data File : VV016703.D
 Acq On : 09 Jun 2020 09:43
 Operator : SY/MD
 Sample : L2862-02
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Jun 10 01:01:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jun 10 00:34:30 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	360172	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	335107	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	155861	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	113235	44.56	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	89.12%
7) Chloroethane-d5	1.56	69	67021	43.95	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	87.90%
11) 1,1-Dichloroethene-d2	2.12	63	150283	35.21	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	70.42%
21) 2-Butanone-d5	3.90	46	181380	96.43	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	96.43%
24) Chloroform-d	4.37	84	236213	45.71	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.42%
26) 1,2-Dichloroethane-d4	5.05	65	169724	48.29	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.58%
32) Benzene-d6	5.07	84	459285	47.55	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.10%
36) 1,2-Dichloropropane-d6	6.09	67	145547	48.34	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.68%
41) Toluene-d8	7.34	98	408954	46.13	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.26%
43) trans-1,3-Dichloropropene-	7.64	79	68954	43.97	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.94%
47) 2-Hexanone-d5	8.11	63	118812	90.50	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	90.50%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	187201	46.96	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	93.92%
64) 1,2-Dichlorobenzene-d4	11.65	152	166743	50.13	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.26%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV060920\
Data File : VV016703.D
Acq On : 09 Jun 2020 09:43
Operator : SY/MD
Sample : L2862-02
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Jun 10 01:01:58 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jun 10 00:34:30 2020
Response via : Initial Calibration

