

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW060922\
 Data File : VW026053.D
 Acq On : 09 Jun 2022 14:33
 Operator : SY/MD
 Sample : N3247-02
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EXYS6

Quant Time: Jun 10 02:52:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR060822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jun 10 02:49:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	177149	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	177120	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	70303	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.314	65	46881	4.508	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	90.200%	
7) Chloroethane-d5	1.574	69	56195	5.169	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	103.400%	
11) 1,1-Dichloroethene-d2	2.114	63	76690	3.193	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	63.800%	
20) 2-Butanone-d5	3.905	46	87879	57.137	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	114.280%	
24) Chloroform-d	4.352	84	118847	5.392	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	107.800%	
26) 1,2-Dichloroethane-d4	5.037	65	51343	5.711	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	114.200%	
32) Benzene-d6	5.050	84	203802	4.762	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	95.200%	
36) 1,2-Dichloropropane-d6	6.069	67	72295	5.460	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	109.200%	
41) Toluene-d8	7.313	98	155621	4.114	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	82.200%	
43) trans-1,3-Dichloroprop...	7.625	79	17590	4.512	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	90.200%	
46) 2-Hexanone-d5	8.088	63	84546	49.283	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	98.560%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	46165	5.171	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	103.400%	
66) 1,2-Dichlorobenzene-d4	11.622	152	61454	5.714	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	114.200%	
Target Compounds						
13) Acetone	2.195	43	13626	11.109	ug/L	79
16) Methylene chloride	2.516	84	3593	0.253	ug/L	95
25) Chloroform	4.388	83	6586	0.262	ug/L	92
29) 1,1,1-Trichloroethane	4.613	97	1659	0.071	ug/L #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV060922\
Data File : VV026053.D
Acq On : 09 Jun 2022 14:33
Operator : SY/MD
Sample : N3247-02
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EXYS6

Quant Time: Jun 10 02:52:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR060822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jun 10 02:49:22 2022
Response via : Initial Calibration

