

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW060922\
 Data File : VW026057.D
 Acq On : 09 Jun 2022 16:09
 Operator : SY/MD
 Sample : N3247-08
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EXYW0

Quant Time: Jun 10 02:52:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR060822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jun 10 02:49:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.625	114	173022	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	170031	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	68352	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.320	65	42563	4.191	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	83.800%	
7) Chloroethane-d5	1.580	69	50654	4.770	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	95.400%	
11) 1,1-Dichloroethene-d2	2.121	63	66525	2.836	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	56.800%#	
20) 2-Butanone-d5	3.908	46	77529	51.610	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	103.220%	
24) Chloroform-d	4.362	84	108178	5.025	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	100.600%	
26) 1,2-Dichloroethane-d4	5.043	65	48290	5.500	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	110.000%	
32) Benzene-d6	5.059	84	189591	4.614	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	92.200%	
36) 1,2-Dichloropropane-d6	6.075	67	67129	5.281	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	105.600%	
41) Toluene-d8	7.320	98	143367	3.949	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	79.000%	
43) trans-1,3-Dichloroprop...	7.628	79	16267	4.347	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	87.000%	
46) 2-Hexanone-d5	8.091	63	79832	48.475	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	96.940%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	43343	5.057	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	101.200%	
66) 1,2-Dichlorobenzene-d4	11.625	152	55969	5.352	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	107.000%	
Target Compounds						
					Qvalue	
14) Carbon disulfide	2.307	76	3423	0.086	ug/L #	91
25) Chloroform	4.384	83	6502	0.265	ug/L	99
29) 1,1,1-Trichloroethane	4.619	97	4060	0.182	ug/L	97
47) Tetrachloroethene	7.982	164	926	0.086	ug/L #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV060922\
Data File : VV026057.D
Acq On : 09 Jun 2022 16:09
Operator : SY/MD
Sample : N3247-08
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EXYW0

Quant Time: Jun 10 02:52:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR060822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jun 10 02:49:22 2022
Response via : Initial Calibration

