

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VV060922\  
 Data File : VV026067.D  
 Acq On : 09 Jun 2022 20:57  
 Operator : SY/MD  
 Sample : VV0609WBL02  
 Misc : 25mL/MSVOA\_V/WATER  
 ALS Vial : 25 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 VBLK365

Quant Time: Jun 10 05:26:18 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR060822WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Fri Jun 10 05:25:55 2022  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon  | Response | Conc     | Units | Dev(Min)  |
|-------------------------------|--------|-------|----------|----------|-------|-----------|
| -----                         |        |       |          |          |       |           |
| Internal Standards            |        |       |          |          |       |           |
| 1) 1,4-Difluorobenzene        | 5.625  | 114   | 153199   | 5.000    | ug/L  | 0.00      |
| 28) Chlorobenzene-d5          | 8.854  | 117   | 144733   | 5.000    | ug/L  | 0.00      |
| 58) 1,4-Dichlorobenzene-d4    | 11.249 | 152   | 60276    | 5.000    | ug/L  | 0.00      |
| System Monitoring Compounds   |        |       |          |          |       |           |
| 4) Vinyl Chloride-d3          | 1.320  | 65    | 41657    | 4.632    | ug/L  | 0.00      |
| Spiked Amount                 | 5.000  | Range | 40 - 130 | Recovery | =     | 92.600%   |
| 7) Chloroethane-d5            | 1.581  | 69    | 49434    | 5.258    | ug/L  | 0.00      |
| Spiked Amount                 | 5.000  | Range | 65 - 130 | Recovery | =     | 105.200%  |
| 11) 1,1-Dichloroethene-d2     | 2.121  | 63    | 65995    | 3.177    | ug/L  | 0.00      |
| Spiked Amount                 | 5.000  | Range | 60 - 125 | Recovery | =     | 63.600%   |
| 20) 2-Butanone-d5             | 3.915  | 46    | 86384    | 64.946   | ug/L  | 0.00      |
| Spiked Amount                 | 50.000 | Range | 40 - 130 | Recovery | =     | 129.900%  |
| 24) Chloroform-d              | 4.362  | 84    | 104846   | 5.501    | ug/L  | 0.00      |
| Spiked Amount                 | 5.000  | Range | 70 - 125 | Recovery | =     | 110.000%  |
| 26) 1,2-Dichloroethane-d4     | 5.043  | 65    | 47325    | 6.087    | ug/L  | 0.00      |
| Spiked Amount                 | 5.000  | Range | 70 - 130 | Recovery | =     | 121.800%  |
| 32) Benzene-d6                | 5.060  | 84    | 186363   | 5.329    | ug/L  | 0.00      |
| Spiked Amount                 | 5.000  | Range | 70 - 125 | Recovery | =     | 106.600%  |
| 36) 1,2-Dichloropropane-d6    | 6.076  | 67    | 65817    | 6.083    | ug/L  | 0.00      |
| Spiked Amount                 | 5.000  | Range | 60 - 140 | Recovery | =     | 121.600%  |
| 41) Toluene-d8                | 7.320  | 98    | 140022   | 4.530    | ug/L  | 0.00      |
| Spiked Amount                 | 5.000  | Range | 70 - 130 | Recovery | =     | 90.600%   |
| 43) trans-1,3-Dichloroprop... | 7.629  | 79    | 15832    | 4.970    | ug/L  | 0.00      |
| Spiked Amount                 | 5.000  | Range | 55 - 130 | Recovery | =     | 99.400%   |
| 46) 2-Hexanone-d5             | 8.092  | 63    | 81934    | 58.447   | ug/L  | 0.00      |
| Spiked Amount                 | 50.000 | Range | 45 - 130 | Recovery | =     | 116.900%  |
| 56) 1,1,2,2-Tetrachloroeth... | 10.217 | 84    | 43639    | 5.982    | ug/L  | 0.00      |
| Spiked Amount                 | 5.000  | Range | 65 - 120 | Recovery | =     | 119.600%  |
| 66) 1,2-Dichlorobenzene-d4    | 11.622 | 152   | 54707    | 5.932    | ug/L  | 0.00      |
| Spiked Amount                 | 5.000  | Range | 80 - 120 | Recovery | =     | 118.600%  |
| Target Compounds              |        |       |          |          |       |           |
| 16) Methylene chloride        | 2.523  | 84    | 4047     | 0.330    | ug/L  | Qvalue 95 |
| -----                         |        |       |          |          |       |           |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VV060922\  
Data File : VV026067.D  
Acq On : 09 Jun 2022 20:57  
Operator : SY/MD  
Sample : VV0609WBL02  
Misc : 25mL/MSVOA\_V/WATER  
ALS Vial : 25 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
VBLK365

Quant Time: Jun 10 05:26:18 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR060822WMA.M  
Quant Title : TRACE VOA SFAM1.0  
QLast Update : Fri Jun 10 05:25:55 2022  
Response via : Initial Calibration

