

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV060922\
 Data File : VV026072.D
 Acq On : 09 Jun 2022 22:56
 Operator : SY/MD
 Sample : N3272-09
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0B53

Quant Time: Jun 10 05:44:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR060822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jun 10 05:25:55 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.629	114	169399	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	163678	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	65159	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.323	65	37481	3.769	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	75.400%	
7) Chloroethane-d5	1.584	69	46619	4.484	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	89.600%	
11) 1,1-Dichloroethene-d2	2.124	63	62938	2.740	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	54.800%#	
20) 2-Butanone-d5	3.912	46	84242	57.279	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	114.560%	
24) Chloroform-d	4.365	84	104873	4.976	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	99.600%	
26) 1,2-Dichloroethane-d4	5.047	65	46345	5.391	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	107.800%	
32) Benzene-d6	5.060	84	183926	4.650	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	93.000%	
36) 1,2-Dichloropropane-d6	6.076	67	65011	5.313	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	106.200%	
41) Toluene-d8	7.320	98	135033	3.863	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	77.200%	
43) trans-1,3-Dichloroprop...	7.632	79	14093	3.912	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	78.200%	
46) 2-Hexanone-d5	8.092	63	77631	48.968	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	97.940%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	44267	5.365	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	107.400%	
66) 1,2-Dichlorobenzene-d4	11.625	152	54869	5.504	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	110.000%	
Target Compounds						
					Qvalue	
9) Trichlorofluoromethane	1.770	101	6346	0.255	ug/L	97
34) Trichloroethene	5.924	95	36190	2.693	ug/L	96
47) Tetrachloroethene	7.979	164	1814	0.174	ug/L	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV060922\
Data File : VV026072.D
Acq On : 09 Jun 2022 22:56
Operator : SY/MD
Sample : N3272-09
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0B53

Quant Time: Jun 10 05:44:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR060822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jun 10 05:25:55 2022
Response via : Initial Calibration

