

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW060922\
 Data File : VW026081.D
 Acq On : 10 Jun 2022 02:31
 Operator : SY/MD
 Sample : N3272-18
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0B63

Quant Time: Jun 10 05:47:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR060822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jun 10 05:25:55 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.625	114	169997	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	164546	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	66254	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.320	65	38179	3.826	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	76.600%	
7) Chloroethane-d5	1.584	69	45893	4.399	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	88.000%	
11) 1,1-Dichloroethene-d2	2.121	63	63053	2.735	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	54.800%#	
20) 2-Butanone-d5	3.921	46	75835	51.381	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	102.760%	
24) Chloroform-d	4.362	84	105581	4.992	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	99.800%	
26) 1,2-Dichloroethane-d4	5.043	65	45741	5.302	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	106.000%	
32) Benzene-d6	5.060	84	180335	4.535	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	90.800%	
36) 1,2-Dichloropropane-d6	6.076	67	65107	5.293	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	105.800%	
41) Toluene-d8	7.320	98	135295	3.850	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	77.000%	
43) trans-1,3-Dichloroprop...	7.632	79	13969	3.857	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	77.200%	
46) 2-Hexanone-d5	8.092	63	78078	48.990	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	97.980%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	44586	5.376	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	107.600%	
66) 1,2-Dichlorobenzene-d4	11.625	152	54225	5.350	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	107.000%	
Target Compounds						
					Qvalue	
9) Trichlorofluoromethane	1.767	101	1251	0.050	ug/L	94
22) cis-1,2-Dichloroethene	3.934	96	2114	0.166	ug/L	79
34) Trichloroethene	5.921	95	92478	6.845	ug/L	98
47) Tetrachloroethene	7.979	164	2752	0.263	ug/L	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV060922\
Data File : VV026081.D
Acq On : 10 Jun 2022 02:31
Operator : SY/MD
Sample : N3272-18
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0B63

Quant Time: Jun 10 05:47:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR060822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jun 10 05:25:55 2022
Response via : Initial Calibration

