

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV060925\
 Data File : VV038769.D
 Acq On : 09 Jun 2025 16:45
 Operator : SY/MD
 Sample : VV0609MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK261

Quant Time: Jun 10 09:10:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM060925WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Jun 10 09:06:39 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	565183	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.776	117	577312	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.175	152	242021	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.288	65	215996	45.954	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	91.900%		
7) Chloroethane-d5	1.542	69	195634	45.471	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	90.940%		
11) 1,1-Dichloroethene-d2	2.069	65	105325	45.528	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	91.060%		
21) 2-Butanone-d5	3.783	46	339019	92.486	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	92.490%		
24) Chloroform-d	4.252	84	455571	45.496	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	91.000%		
26) 1,2-Dichloroethane-d4	4.941	65	287942	49.256	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	98.520%		
32) Benzene-d6	4.960	84	874303	47.999	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	96.000%		
36) 1,2-Dichloropropane-d6	5.986	67	298187	49.290	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	98.580%		
41) Toluene-d8	7.239	98	736996	45.820	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	91.640%		
43) trans-1,3-Dichloroprop...	7.548	79	119928	45.457	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	90.920%		
47) 2-Hexanone-d5	8.014	63	253867	101.146	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	101.150%		
56) 1,1,2,2-Tetrachloroeth...	10.143	84	410382	45.718	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	91.440%		
66) 1,2-Dichlorobenzene-d4	11.548	152	257940	54.915	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	109.840%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV060925\
Data File : VV038769.D
Acq On : 09 Jun 2025 16:45
Operator : SY/MD
Sample : VV0609MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK261

Quant Time: Jun 10 09:10:37 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM060925WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Jun 10 09:06:39 2025
Response via : Initial Calibration

