

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061019\
 Data File : VV011376.D
 Acq On : 11 Jun 2019 05:32
 Operator : SY/MD
 Sample : K3229-11
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F9P31

Quant Time: Jun 11 08:29:22 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Jun 11 08:01:21 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	138730	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	127934	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	49791	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	37203	3.54	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	70.80%
7) Chloroethane-d5	1.57	69	34939	4.00	ug/L	0.02
Spiked Amount	5.000	Range	65 - 130	Recovery	=	80.00%
11) 1,1-Dichloroethene-d2	2.12	63	63989	3.05	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	61.00%
20) 2-Butanone-d5	3.96	46	140508	59.37	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	118.74%
24) Chloroform-d	4.40	84	83580	4.52	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.40%
26) 1,2-Dichloroethane-d4	5.08	65	50643	5.06	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.20%
32) Benzene-d6	5.09	84	164142	4.29	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.80%
36) 1,2-Dichloropropane-d6	6.12	67	52196	4.43	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	88.60%
41) Toluene-d8	7.36	98	136064	3.87	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	77.40%
43) trans-1,3-Dichloropropene-	7.67	79	14184	3.16	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	63.20%
46) 2-Hexanone-d5	8.14	63	102174	53.82	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	107.64%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	39526	4.92	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.40%
64) 1,2-Dichlorobenzene-d4	11.67	152	46080	4.69	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.80%

Target Compounds					Ovalue
13) Acetone	2.21	43	10211	6.435 ug/L	83
14) Carbon disulfide	2.31	76	7448	0.281 ug/L #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV061019\
Data File : VV011376.D
Acq On : 11 Jun 2019 05:32
Operator : SY/MD
Sample : K3229-11
Misc : 25ML/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F9P31

Quant Time: Jun 11 08:29:22 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jun 11 08:01:21 2019
Response via : Initial Calibration

