

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061019\
 Data File : VV011383.D
 Acq On : 11 Jun 2019 08:49
 Operator : SY/MD
 Sample : K3226-16
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFCPO

Quant Time: Jun 12 02:39:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Jun 12 02:23:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	142326	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	134568	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	52711	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	37146	3.44	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	68.80%
7) Chloroethane-d5	1.55	69	33798	3.77	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	75.40%
11) 1,1-Dichloroethene-d2	2.12	63	60945	2.83	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	56.60%#
20) 2-Butanone-d5	3.94	46	140638	57.92	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	115.84%
24) Chloroform-d	4.40	84	83376	4.39	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.80%
26) 1,2-Dichloroethane-d4	5.08	65	49261	4.80	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.00%
32) Benzene-d6	5.09	84	162453	4.04	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	80.80%
36) 1,2-Dichloropropane-d6	6.12	67	51388	4.14	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	82.80%
41) Toluene-d8	7.36	98	140351	3.80	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	76.00%
43) trans-1,3-Dichloropropene-	7.66	79	14512	3.07	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	61.40%
46) 2-Hexanone-d5	8.13	63	105151	52.65	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	105.30%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	40103	4.74	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.80%
64) 1,2-Dichlorobenzene-d4	11.67	152	48768	4.69	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.80%

Target Compounds					Ovalue
22) cis-1,2-Dichloroethene	3.96	96	9216	0.829 ug/L	93
34) Trichloroethene	5.96	95	251234	22.711 ug/L	98
42) Toluene	7.43	91	7505	0.164 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV061019\
Data File : VV011383.D
Acq On : 11 Jun 2019 08:49
Operator : SY/MD
Sample : K3226-16
Misc : 25ML/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFCP0

Quant Time: Jun 12 02:39:23 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Jun 12 02:23:01 2019
Response via : Initial Calibration

