

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061019\
 Data File : VV011396.D
 Acq On : 11 Jun 2019 16:24
 Operator : SY/MD
 Sample : VIBLK51
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK51

Quant Time: Jun 12 04:08:51 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Jun 12 02:48:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	128248	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	116841	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	45623	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	31515	3.24	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	64.80%
7) Chloroethane-d5	1.56	69	30941	3.83	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	76.60%
11) 1,1-Dichloroethene-d2	2.12	63	53106	2.74	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	54.80%#
20) 2-Butanone-d5	3.95	46	113098	51.69	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.38%
24) Chloroform-d	4.40	84	74467	4.36	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.20%
26) 1,2-Dichloroethane-d4	5.08	65	43271	4.68	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.60%
32) Benzene-d6	5.09	84	141263	4.04	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	80.80%
36) 1,2-Dichloropropane-d6	6.12	67	46359	4.31	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	86.20%
41) Toluene-d8	7.36	98	117573	3.66	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	73.20%
43) trans-1,3-Dichloropropene-	7.67	79	12455	3.03	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	60.60%
46) 2-Hexanone-d5	8.13	63	84854	48.94	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	97.88%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	33171	4.52	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.40%
64) 1,2-Dichlorobenzene-d4	11.67	152	41944	4.66	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV061019\
Data File : VV011396.D
Acq On : 11 Jun 2019 16:24
Operator : SY/MD
Sample : VIBLK51
Misc : 25ML/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK51

Quant Time: Jun 12 04:08:51 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Jun 12 02:48:56 2019
Response via : Initial Calibration

