

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061019\
 Data File : VV011400.D
 Acq On : 11 Jun 2019 18:55
 Operator : SY/MD
 Sample : VIBLK53
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK53

Quant Time: Jun 12 04:24:21 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Jun 12 02:48:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	139435	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	128340	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	50752	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	32398	3.07	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	61.40%
7) Chloroethane-d5	1.56	69	32824	3.74	ug/L	0.01
Spiked Amount	5.000	Range	65 - 130	Recovery	=	74.80%
11) 1,1-Dichloroethene-d2	2.12	63	55499	2.63	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	52.60%#
20) 2-Butanone-d5	3.95	46	130312	54.78	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	109.56%
24) Chloroform-d	4.40	84	78924	4.25	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.00%
26) 1,2-Dichloroethane-d4	5.08	65	47671	4.74	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.80%
32) Benzene-d6	5.09	84	148608	3.87	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	77.40%
36) 1,2-Dichloropropane-d6	6.12	67	49432	4.18	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	83.60%
41) Toluene-d8	7.36	98	123754	3.51	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	70.20%
43) trans-1,3-Dichloropropene-	7.66	79	13061	2.90	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	58.00%
46) 2-Hexanone-d5	8.14	63	97227	51.05	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	102.10%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	36320	4.50	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.00%
64) 1,2-Dichlorobenzene-d4	11.67	152	42962	4.29	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	85.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV061019\
Data File : VV011400.D
Acq On : 11 Jun 2019 18:55
Operator : SY/MD
Sample : VIBLK53
Misc : 25ML/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK53

Quant Time: Jun 12 04:24:21 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Jun 12 02:48:56 2019
Response via : Initial Calibration

