

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061019\
 Data File : VV011402.D
 Acq On : 11 Jun 2019 19:52
 Operator : SY/MD
 Sample : K3291-02
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DB8G1

Quant Time: Jun 12 04:31:24 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Jun 12 02:48:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	128767	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	125429	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	50247	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	24608	2.52	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	50.40%
7) Chloroethane-d5	1.55	69	28055	3.46	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	69.20%
11) 1,1-Dichloroethene-d2	2.11	63	43511	2.23	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	44.60%#
20) 2-Butanone-d5	3.94	46	135226	61.56	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	123.12%
24) Chloroform-d	4.40	84	70614	4.11	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.20%
26) 1,2-Dichloroethane-d4	5.08	65	45536	4.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.00%
32) Benzene-d6	5.09	84	128929	3.44	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	68.80%#
36) 1,2-Dichloropropane-d6	6.11	67	46405	4.02	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	80.40%
41) Toluene-d8	7.36	98	108489	3.15	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	63.00%#
43) trans-1,3-Dichloropropene-	7.66	79	13143	2.98	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	59.60%
46) 2-Hexanone-d5	8.13	63	105103	56.47	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	112.94%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	39645	5.03	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.60%
64) 1,2-Dichlorobenzene-d4	11.67	152	42401	4.28	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	85.60%

Target Compounds					Ovalue
13) Acetone	2.19	43	2766	1.878 ug/L	84
22) cis-1,2-Dichloroethene	3.95	96	27403	2.724 ug/L	95
34) Trichloroethene	5.95	95	669544	64.937 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061019\
Data File : VV011402.D
Acq On : 11 Jun 2019 19:52
Operator : SY/MD
Sample : K3291-02
Misc : 25ML/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DB8G1

Quant Time: Jun 12 04:31:24 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Jun 12 02:48:56 2019
Response via : Initial Calibration

