

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061019\
 Data File : VV011407.D
 Acq On : 11 Jun 2019 23:39
 Operator : SY/MD
 Sample : K3289-01
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFCL3

Quant Time: Jun 12 04:53:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Jun 12 04:39:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	133900	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	124994	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	48750	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	29695	2.93	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	58.60%
7) Chloroethane-d5	1.55	69	31519	3.74	ug/L	-0.02
Spiked Amount	5.000	Range	65 - 130	Recovery	=	74.80%
11) 1,1-Dichloroethene-d2	2.12	63	51609	2.55	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	51.00%#
20) 2-Butanone-d5	3.94	46	140215	61.38	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	122.76%
24) Chloroform-d	4.40	84	77527	4.34	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.80%
26) 1,2-Dichloroethane-d4	5.08	65	45869	4.75	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.00%
32) Benzene-d6	5.09	84	143792	3.85	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	77.00%
36) 1,2-Dichloropropane-d6	6.12	67	48593	4.22	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	84.40%
41) Toluene-d8	7.36	98	120165	3.50	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	70.00%
43) trans-1,3-Dichloropropene-	7.66	79	12944	2.95	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	59.00%
46) 2-Hexanone-d5	8.13	63	102615	55.32	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	110.64%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	38226	4.87	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.40%
64) 1,2-Dichlorobenzene-d4	11.67	152	44243	4.60	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	92.00%

Target Compounds					Ovalue
13) Acetone	2.19	43	4493	2.934 ug/L	96
19) 1,1-Dichloroethane	3.22	63	2992	0.161 ug/L	95
22) cis-1,2-Dichloroethene	3.95	96	2672	0.255 ug/L	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061019\
Data File : VV011407.D
Acq On : 11 Jun 2019 23:39
Operator : SY/MD
Sample : K3289-01
Misc : 25ML/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFCL3

Quant Time: Jun 12 04:53:09 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Jun 12 04:39:23 2019
Response via : Initial Calibration

