

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061019\
 Data File : VV011362.D
 Acq On : 10 Jun 2019 20:50
 Operator : SY/MD
 Sample : VIBLK48
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK48

Quant Time: Jun 11 03:46:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Jun 11 01:32:35 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	155961	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	144863	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	55548	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	43154	3.65	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	73.00%
7) Chloroethane-d5	1.55	69	39194	3.99	ug/L	-0.02
Spiked Amount	5.000	Range	65 - 130	Recovery	=	79.80%
11) 1,1-Dichloroethene-d2	2.12	63	69828	2.96	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	59.20%#
20) 2-Butanone-d5	3.94	46	132539	49.82	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	99.64%
24) Chloroform-d	4.40	84	90117	4.33	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.60%
26) 1,2-Dichloroethane-d4	5.08	65	52799	4.70	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.00%
32) Benzene-d6	5.09	84	178354	4.12	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.40%
36) 1,2-Dichloropropane-d6	6.12	67	57207	4.29	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	85.80%
41) Toluene-d8	7.36	98	155640	3.91	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.20%
43) trans-1,3-Dichloropropene-	7.66	79	16280	3.20	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	64.00%
46) 2-Hexanone-d5	8.13	63	100280	46.65	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.30%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	40558	4.46	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	89.20%
64) 1,2-Dichlorobenzene-d4	11.67	152	50331	4.59	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	91.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV061019\
Data File : VV011362.D
Acq On : 10 Jun 2019 20:50
Operator : SY/MD
Sample : VIBLK48
Misc : 25ML/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK48

Quant Time: Jun 11 03:46:55 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jun 11 01:32:35 2019
Response via : Initial Calibration

