

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061019\
 Data File : VV011393.D
 Acq On : 11 Jun 2019 14:30
 Operator : SY/MD
 Sample : K3197-16
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Jun 12 03:53:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Jun 12 02:48:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	139866	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	126554	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	50292	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	36541	3.45	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	69.00%
7) Chloroethane-d5	1.56	69	34791	3.95	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	79.00%
11) 1,1-Dichloroethene-d2	2.12	63	63397	2.99	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	59.80%#
20) 2-Butanone-d5	3.95	46	117020	49.04	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	98.08%
24) Chloroform-d	4.40	84	83123	4.46	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.20%
26) 1,2-Dichloroethane-d4	5.08	65	46679	4.63	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.60%
32) Benzene-d6	5.09	84	165900	4.38	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.60%
36) 1,2-Dichloropropane-d6	6.12	67	52444	4.50	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	90.00%
41) Toluene-d8	7.36	98	141242	4.06	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	81.20%
43) trans-1,3-Dichloropropene-	7.66	79	14091	3.17	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	63.40%
46) 2-Hexanone-d5	8.13	63	86212	45.90	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	91.80%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	34757	4.37	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	87.40%
64) 1,2-Dichlorobenzene-d4	11.67	152	47201	4.75	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061019\
Data File : VV011393.D
Acq On : 11 Jun 2019 14:30
Operator : SY/MD
Sample : K3197-16
Misc : 25ML/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Jun 12 03:53:09 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR060419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Jun 12 02:48:56 2019
Response via : Initial Calibration

