

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061020\
 Data File : VV016763.D
 Acq On : 10 Jun 2020 12:52
 Operator : SY/MD
 Sample : L2861-21
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Jun 11 01:42:43 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jun 11 01:36:13 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	352304	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	337355	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	166469	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	115984	46.66	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	93.32%
7) Chloroethane-d5	1.58	69	81593	54.70	ug/L	0.02
Spiked Amount	50.000	Range	70 - 130	Recovery	=	109.40%
11) 1,1-Dichloroethene-d2	2.12	63	154565	37.03	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	74.06%
21) 2-Butanone-d5	3.89	46	185468	100.80	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	100.80%
24) Chloroform-d	4.37	84	240518	47.58	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.16%
26) 1,2-Dichloroethane-d4	5.05	65	172024	50.03	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.06%
32) Benzene-d6	5.07	84	469245	48.25	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.50%
36) 1,2-Dichloropropane-d6	6.09	67	147499	48.66	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.32%
41) Toluene-d8	7.33	98	429680	48.15	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.30%
43) trans-1,3-Dichloropropene-	7.64	79	75561	47.86	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	95.72%
47) 2-Hexanone-d5	8.10	63	111010	84.00	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	84.00%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	191382	47.69	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	95.38%
64) 1,2-Dichlorobenzene-d4	11.65	152	180129	50.70	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061020\
Data File : VV016763.D
Acq On : 10 Jun 2020 12:52
Operator : SY/MD
Sample : L2861-21
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Jun 11 01:42:43 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jun 11 01:36:13 2020
Response via : Initial Calibration

