

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061020\
 Data File : VV016767.D
 Acq On : 10 Jun 2020 14:27
 Operator : SY/MD
 Sample : L2901-10 100X
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 ETTT5

Quant Time: Jun 11 02:05:18 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jun 11 01:04:46 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	362135	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	343078	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	172213	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	114550	44.83	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	89.66%
7) Chloroethane-d5	1.56	69	76307	49.77	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	99.54%
11) 1,1-Dichloroethene-d2	2.12	63	152652	35.58	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.16%
21) 2-Butanone-d5	3.89	46	172935	91.44	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	91.44%
24) Chloroform-d	4.37	84	233121	44.87	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.74%
26) 1,2-Dichloroethane-d4	5.05	65	167565	47.41	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.82%
32) Benzene-d6	5.07	84	461275	46.64	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.28%
36) 1,2-Dichloropropane-d6	6.09	67	143008	46.39	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	92.78%
41) Toluene-d8	7.34	98	422096	46.51	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.02%
43) trans-1,3-Dichloropropene-	7.64	79	72340	45.06	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	90.12%
47) 2-Hexanone-d5	8.11	63	106675	79.37	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	79.37%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	182438	44.70	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.40%
64) 1,2-Dichlorobenzene-d4	11.65	152	178962	48.69	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.38%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061020\
Data File : VV016767.D
Acq On : 10 Jun 2020 14:27
Operator : SY/MD
Sample : L2901-10 100X
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
ETTT5

Quant Time: Jun 11 02:05:18 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jun 11 01:04:46 2020
Response via : Initial Calibration

