

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061020\  
 Data File : VV016775.D  
 Acq On : 10 Jun 2020 17:38  
 Operator : SY/MD  
 Sample : L2914-06 20X  
 Misc : 5.0mL/MSVOA V/WATER  
 ALS Vial : 16 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 F9E01

Manual Integrations  
 APPROVED

MMDadoda  
 6/11/2020 4:23:26 PM

Quant Time: Jun 11 02:49:20 2020  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_V\METHOD\SOMVLM060820WMA.M  
 Quant Title : VOC Analysis  
 QLast Update : Thu Jun 11 01:04:46 2020  
 Response via : Initial Calibration

| Internal Standards         | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|----------------------------|-------|------|----------|-------|-------|----------|
| 1) 1,4-Difluorobenzene     | 5.64  | 114  | 355187   | 50.00 | ug/L  | 0.00     |
| 28) Chlorobenzene-d5       | 8.87  | 117  | 334840   | 50.00 | ug/L  | 0.00     |
| 60) 1,4-Dichlorobenzene-d4 | 11.27 | 152  | 176291   | 50.00 | ug/L  | 0.00     |

## System Monitoring Compounds

|                                |         |       |          |          |      |        |
|--------------------------------|---------|-------|----------|----------|------|--------|
| 4) Vinyl Chloride-d3           | 1.31    | 65    | 118588   | 47.32    | ug/L | 0.00   |
| Spiked Amount                  | 50.000  | Range | 60 - 135 | Recovery | =    | 94.64% |
| 7) Chloroethane-d5             | 1.54    | 69    | 59848    | 39.80    | ug/L | -0.02  |
| Spiked Amount                  | 50.000  | Range | 70 - 130 | Recovery | =    | 79.60% |
| 11) 1,1-Dichloroethene-d2      | 2.11    | 63    | 140521   | 33.39    | ug/L | 0.00   |
| Spiked Amount                  | 50.000  | Range | 60 - 125 | Recovery | =    | 66.78% |
| 21) 2-Butanone-d5              | 3.90    | 46    | 164803   | 88.84    | ug/L | 0.00   |
| Spiked Amount                  | 100.000 | Range | 40 - 130 | Recovery | =    | 88.84% |
| 24) Chloroform-d               | 4.37    | 84    | 212273   | 41.65    | ug/L | 0.00   |
| Spiked Amount                  | 50.000  | Range | 70 - 125 | Recovery | =    | 83.30% |
| 26) 1,2-Dichloroethane-d4      | 5.05    | 65    | 157486   | 45.43    | ug/L | 0.00   |
| Spiked Amount                  | 50.000  | Range | 70 - 125 | Recovery | =    | 90.86% |
| 32) Benzene-d6                 | 5.07    | 84    | 436784   | 45.25    | ug/L | 0.00   |
| Spiked Amount                  | 50.000  | Range | 70 - 125 | Recovery | =    | 90.50% |
| 36) 1,2-Dichloropropane-d6     | 6.09    | 67    | 133307   | 44.31    | ug/L | 0.00   |
| Spiked Amount                  | 50.000  | Range | 70 - 120 | Recovery | =    | 88.62% |
| 41) Toluene-d8                 | 7.34    | 98    | 392984   | 44.37    | ug/L | 0.00   |
| Spiked Amount                  | 50.000  | Range | 80 - 120 | Recovery | =    | 88.74% |
| 43) trans-1,3-Dichloropropene- | 7.64    | 79    | 65395    | 41.73    | ug/L | 0.00   |
| Spiked Amount                  | 50.000  | Range | 60 - 125 | Recovery | =    | 83.46% |
| 47) 2-Hexanone-d5              | 8.11    | 63    | 101911   | 77.69    | ug/L | 0.00   |
| Spiked Amount                  | 100.000 | Range | 45 - 130 | Recovery | =    | 77.69% |
| 57) 1,1,2,2-Tetrachloroethane- | 10.24   | 84    | 181126   | 45.47    | ug/L | 0.00   |
| Spiked Amount                  | 50.000  | Range | 65 - 120 | Recovery | =    | 90.94% |
| 64) 1,2-Dichlorobenzene-d4     | 11.65   | 152   | 174940   | 46.50    | ug/L | 0.00   |
| Spiked Amount                  | 50.000  | Range | 80 - 120 | Recovery | =    | 93.00% |

## Target Compounds

| Target Compounds      | R.T. | QIon | Response | Conc    | Units  | Ovalue |
|-----------------------|------|------|----------|---------|--------|--------|
| 25) Chloroform        | 4.39 | 83   | 7592m    | 1.633   | ug/L   |        |
| 29) Cyclohexane       | 4.69 | 56   | 15302    | 3.678   | ug/L   | 97     |
| 33) Benzene           | 5.12 | 78   | 1982557  | 200.118 | ug/L   | 100    |
| 35) Methylcyclohexane | 6.15 | 83   | 2586     | 0.620   | ug/L # | 63     |
| 42) Toluene           | 7.41 | 91   | 41623    | 3.985   | ug/L   | 96     |
| 52) Ethylbenzene      | 9.04 | 91   | 30586    | 2.644   | ug/L   | 97     |
| 53) m,p-Xylene        | 9.16 | 106  | 28297    | 6.570   | ug/L   | 98     |
| 54) o-xylene          | 9.57 | 106  | 10767    | 2.535   | ug/L   | 91     |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV061020\  
Data File : VV016775.D  
Acq On : 10 Jun 2020 17:38  
Operator : SY/MD  
Sample : L2914-06 20X  
Misc : 5.0mL/MSVOA V/WATER  
ALS Vial : 16 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
F9E01

Quant Time: Jun 11 02:49:20 2020  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_V\METHOD\SOMVLM060820WMA.M  
Quant Title : VOC Analysis  
QLast Update : Thu Jun 11 01:04:46 2020  
Response via : Initial Calibration

Manual Integrations  
APPROVED

MMDadoda  
6/11/2020 4:23:26 PM

