

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061020\
 Data File : VV016784.D
 Acq On : 10 Jun 2020 21:36
 Operator : SY/MD
 Sample : VV0610WBL02
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK79

Quant Time: Jun 11 04:22:34 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jun 11 04:06:22 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	367842	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	347070	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	175198	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	113775	43.83	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	87.66%
7) Chloroethane-d5	1.56	69	77218	49.58	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	99.16%
11) 1,1-Dichloroethene-d2	2.11	63	148941	34.17	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	68.34%
21) 2-Butanone-d5	3.89	46	172321	89.70	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	89.70%
24) Chloroform-d	4.37	84	230737	43.72	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.44%
26) 1,2-Dichloroethane-d4	5.05	65	166099	46.27	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.54%
32) Benzene-d6	5.07	84	451014	45.08	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.16%
36) 1,2-Dichloropropane-d6	6.09	67	139389	44.70	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	89.40%
41) Toluene-d8	7.34	98	416189	45.33	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.66%
43) trans-1,3-Dichloropropene-	7.64	79	68308	42.05	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	84.10%
47) 2-Hexanone-d5	8.11	63	106325	78.20	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	78.20%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	182583	44.23	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	88.46%
64) 1,2-Dichlorobenzene-d4	11.65	152	176577	47.23	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.46%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061020\
Data File : VV016784.D
Acq On : 10 Jun 2020 21:36
Operator : SY/MD
Sample : VV0610WBL02
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK79

Quant Time: Jun 11 04:22:34 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jun 11 04:06:22 2020
Response via : Initial Calibration

