

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061020\
 Data File : VV016790.D
 Acq On : 10 Jun 2020 23:59
 Operator : SY/MD
 Sample : L2929-04 100X
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 ETTW2

Quant Time: Jun 11 05:21:37 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jun 11 04:06:22 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	314249	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	298495	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	151710	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	98964	44.63	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	89.26%
7) Chloroethane-d5	1.55	69	66313	49.84	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	99.68%
11) 1,1-Dichloroethene-d2	2.11	63	130969	35.17	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	70.34%
21) 2-Butanone-d5	3.90	46	159614	97.26	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	97.26%
24) Chloroform-d	4.37	84	202703	44.96	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.92%
26) 1,2-Dichloroethane-d4	5.05	65	147145	47.98	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.96%
32) Benzene-d6	5.07	84	400616	46.56	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.12%
36) 1,2-Dichloropropane-d6	6.09	67	124702	46.49	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	92.98%
41) Toluene-d8	7.34	98	361510	45.78	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.56%
43) trans-1,3-Dichloropropene-	7.64	79	58970	42.21	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	84.42%
47) 2-Hexanone-d5	8.11	63	94665	80.96	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	80.96%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	163715	46.11	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.22%
64) 1,2-Dichlorobenzene-d4	11.65	152	157089	48.52	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.04%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV061020\
Data File : VV016790.D
Acq On : 10 Jun 2020 23:59
Operator : SY/MD
Sample : L2929-04 100X
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
ETTW2

Quant Time: Jun 11 05:21:37 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jun 11 04:06:22 2020
Response via : Initial Calibration

