

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061020\
 Data File : VV016791.D
 Acq On : 11 Jun 2020 00:23
 Operator : SY/MD
 Sample : L2929-05 100X
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 ETTW3

Quant Time: Jun 11 05:29:08 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jun 11 04:06:22 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	319512	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	304589	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	150359	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	107506	47.68	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	95.36%
7) Chloroethane-d5	1.55	69	66634	49.26	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	98.52%
11) 1,1-Dichloroethene-d2	2.11	63	134814	35.61	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.22%
21) 2-Butanone-d5	3.89	46	164058	98.32	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	98.32%
24) Chloroform-d	4.37	84	211307	46.09	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.18%
26) 1,2-Dichloroethane-d4	5.05	65	152880	49.03	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.06%
32) Benzene-d6	5.07	84	413576	47.10	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.20%
36) 1,2-Dichloropropane-d6	6.09	67	126507	46.22	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	92.44%
41) Toluene-d8	7.34	98	372354	46.21	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.42%
43) trans-1,3-Dichloropropene-	7.64	79	62704	43.99	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.98%
47) 2-Hexanone-d5	8.11	63	98582	82.62	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	82.62%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	166526	45.96	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.92%
64) 1,2-Dichlorobenzene-d4	11.65	152	161940	50.47	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.94%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV061020\
Data File : VV016791.D
Acq On : 11 Jun 2020 00:23
Operator : SY/MD
Sample : L2929-05 100X
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
ETTW3

Quant Time: Jun 11 05:29:08 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jun 11 04:06:22 2020
Response via : Initial Calibration

