

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061021\
 Data File : VV021538.D
 Acq On : 10 Jun 2021 12:16
 Operator : SY/MD
 Sample : M2649-11
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 B0AA7

Quant Time: Jun 11 02:44:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM052721WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jun 11 02:43:15 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	445734	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	423844	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	246414	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	147119	50.131	ug/L	0.00
Spiked Amount 50.000	Range 60	- 135	Recovery	=	100.260%	
7) Chloroethane-d5	1.568	69	112991	51.759	ug/L	0.00
Spiked Amount 50.000	Range 70	- 130	Recovery	=	103.520%	
11) 1,1-Dichloroethene-d2	2.108	63	191812	38.251	ug/L	0.00
Spiked Amount 50.000	Range 60	- 125	Recovery	=	76.500%	
21) 2-Butanone-d5	3.889	46	226939	107.841	ug/L	0.00
Spiked Amount 100.000	Range 40	- 130	Recovery	=	107.840%	
24) Chloroform-d	4.352	84	286119	48.977	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	97.960%	
26) 1,2-Dichloroethane-d4	5.034	65	173023	51.421	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	102.840%	
32) Benzene-d6	5.053	84	588926	49.669	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	99.340%	
36) 1,2-Dichloropropane-d6	6.072	67	179687	49.072	ug/L	0.00
Spiked Amount 50.000	Range 70	- 120	Recovery	=	98.140%	
41) Toluene-d8	7.317	98	561359	49.693	ug/L	0.00
Spiked Amount 50.000	Range 80	- 120	Recovery	=	99.380%	
43) trans-1,3-Dichloroprop...	7.622	79	94835	48.396	ug/L	0.00
Spiked Amount 50.000	Range 60	- 125	Recovery	=	96.800%	
47) 2-Hexanone-d5	8.088	63	173592	103.244	ug/L	0.00
Spiked Amount 100.000	Range 45	- 130	Recovery	=	103.240%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	256220	50.409	ug/L	0.00
Spiked Amount 50.000	Range 65	- 120	Recovery	=	100.820%	
66) 1,2-Dichlorobenzene-d4	11.625	152	252488	49.292	ug/L	0.00
Spiked Amount 50.000	Range 80	- 120	Recovery	=	98.580%	
Target Compounds						Qvalue
13) Acetone	2.185	43	6231	3.441	ug/L	# 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061021\
Data File : VV021538.D
Acq On : 10 Jun 2021 12:16
Operator : SY/MD
Sample : M2649-11
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
B0AA7

Quant Time: Jun 11 02:44:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM052721WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jun 11 02:43:15 2021
Response via : Initial Calibration

