

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061021\
 Data File : VV021540.D
 Acq On : 10 Jun 2021 13:03
 Operator : SY/MD
 Sample : M2662-07
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BGDM1

Quant Time: Jun 11 02:45:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM052721WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jun 11 02:43:15 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	474115	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	445481	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	260714	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	163396	52.345	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 104.700%		
7) Chloroethane-d5	1.568	69	124318	53.538	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	= 107.080%		
11) 1,1-Dichloroethene-d2	2.108	63	216358	40.563	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 81.120%		
21) 2-Butanone-d5	3.889	46	230980	103.191	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	= 103.190%		
24) Chloroform-d	4.352	84	314260	50.574	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 101.140%		
26) 1,2-Dichloroethane-d4	5.034	65	190193	53.140	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 106.280%		
32) Benzene-d6	5.053	84	652744	52.377	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 104.760%		
36) 1,2-Dichloropropane-d6	6.072	67	201152	52.266	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 104.540%		
41) Toluene-d8	7.317	98	625485	52.680	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 105.360%		
43) trans-1,3-Dichloroprop...	7.622	79	104549	50.762	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 101.520%		
47) 2-Hexanone-d5	8.088	63	177859	100.644	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	= 100.640%		
56) 1,1,2,2-Tetrachloroeth...	10.217	84	273780	51.247	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 102.500%		
66) 1,2-Dichlorobenzene-d4	11.625	152	284572	52.508	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 105.020%		
Target Compounds						Qvalue
13) Acetone	2.185	43	16049	8.333	ug/L #	84
25) Chloroform	4.381	83	4094	0.703	ug/L	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061021\
Data File : VV021540.D
Acq On : 10 Jun 2021 13:03
Operator : SY/MD
Sample : M2662-07
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BGDM1

Quant Time: Jun 11 02:45:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM052721WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jun 11 02:43:15 2021
Response via : Initial Calibration

