

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW061022\
 Data File : VW026113.D
 Acq On : 10 Jun 2022 23:17
 Operator : SY/MD
 Sample : N3273-02
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0B71

Quant Time: Jun 11 01:30:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR060822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Jun 11 01:25:15 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.625	114	158047	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	157419	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	65345	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.320	65	50043	5.394	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 107.800%		
7) Chloroethane-d5	1.584	69	56459	5.821	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 116.400%		
11) 1,1-Dichloroethene-d2	2.121	63	82304	3.841	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 76.800%		
20) 2-Butanone-d5	3.931	46	95791	69.809	ug/L	0.03
Spiked Amount	50.000	Range 40 - 130	Recovery	= 139.620%#		
24) Chloroform-d	4.365	84	109615	5.575	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 111.400%		
26) 1,2-Dichloroethane-d4	5.047	65	47836	5.964	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 119.200%		
32) Benzene-d6	5.060	84	203720	5.356	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 107.200%		
36) 1,2-Dichloropropane-d6	6.076	67	66861	5.681	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 113.600%		
41) Toluene-d8	7.320	98	158816	4.724	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 94.400%		
43) trans-1,3-Dichloroprop...	7.632	79	14495	4.184	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 83.600%		
46) 2-Hexanone-d5	8.092	63	73174	47.992	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 95.980%		
56) 1,1,2,2-Tetrachloroeth...	10.217	84	42021	5.296	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 106.000%		
66) 1,2-Dichlorobenzene-d4	11.625	152	56057	5.607	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 112.200%		
Target Compounds						
34) Trichloroethene	5.931	95	2666	0.206	ug/L	Qvalue 90
42) Toluene	7.400	91	10768	0.208	ug/L	95
52) Ethylbenzene	9.021	91	3303	0.061	ug/L	92
53) m,p-Xylene	9.143	106	5382	0.257	ug/L	97
54) o-Xylene	9.548	106	1629	0.082	ug/L	86
63) 1,2,4-Trimethylbenzene	10.918	105	5625	0.164	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061022\
Data File : VV026113.D
Acq On : 10 Jun 2022 23:17
Operator : SY/MD
Sample : N3273-02
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0B71

Quant Time: Jun 11 01:30:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR060822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jun 11 01:25:15 2022
Response via : Initial Calibration

