

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV061120\
 Data File : VV016827.D
 Acq On : 11 Jun 2020 15:19
 Operator : SY/MD
 Sample : L2929-07
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 ETTW6

Quant Time: Jun 12 03:19:44 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jun 12 01:27:40 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	352570	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	334925	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	168715	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	86065	34.59	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	69.18%
7) Chloroethane-d5	1.55	69	61874	41.45	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	82.90%
11) 1,1-Dichloroethene-d2	2.11	63	123282	29.51	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	59.02%#
21) 2-Butanone-d5	3.89	46	162478	88.24	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	88.24%
24) Chloroform-d	4.37	84	200310	39.60	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	79.20%
26) 1,2-Dichloroethane-d4	5.05	65	145714	42.35	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.70%
32) Benzene-d6	5.07	84	382316	39.60	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	79.20%
36) 1,2-Dichloropropane-d6	6.09	67	120931	40.18	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	80.36%
41) Toluene-d8	7.34	98	349354	39.43	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	78.86%#
43) trans-1,3-Dichloropropene-	7.64	79	59153	37.74	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	75.48%
47) 2-Hexanone-d5	8.11	63	101390	77.28	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	77.28%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	168328	42.25	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	84.50%
64) 1,2-Dichlorobenzene-d4	11.65	152	155736	43.25	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	86.50%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_V\DATA\VV061120\
Data File : VV016827.D
Acq On : 11 Jun 2020 15:19
Operator : SY/MD
Sample : L2929-07
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
ETTW6

Quant Time: Jun 12 03:19:44 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jun 12 01:27:40 2020
Response via : Initial Calibration

