

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV061120\
 Data File : VV016856.D
 Acq On : 12 Jun 2020 03:09
 Operator : SY/MD
 Sample : L2915-01DL 250X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F9E04DL

Quant Time: Jun 12 07:46:59 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jun 12 05:55:11 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	369996	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	346018	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	161012	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	107691	41.25	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	82.50%
7) Chloroethane-d5	1.56	69	74679	47.67	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	95.34%
11) 1,1-Dichloroethene-d2	2.12	63	148379	33.85	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	67.70%
21) 2-Butanone-d5	3.90	46	165961	85.89	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	85.89%
24) Chloroform-d	4.37	84	222603	41.93	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.86%
26) 1,2-Dichloroethane-d4	5.05	65	165114	45.73	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.46%
32) Benzene-d6	5.07	84	446470	44.76	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.52%
36) 1,2-Dichloropropane-d6	6.09	67	137767	44.31	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	88.62%
41) Toluene-d8	7.34	98	402694	44.00	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.00%
43) trans-1,3-Dichloropropene-	7.64	79	66310	40.95	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.90%
47) 2-Hexanone-d5	8.11	63	98889	72.95	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	72.95%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	170440	41.41	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	82.82%
64) 1,2-Dichlorobenzene-d4	11.65	152	162251	47.22	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.44%

Target Compounds

					Qvalue
29) Cyclohexane	4.70	56	4422	1.028	ug/L 96
33) Benzene	5.12	78	1434489	140.119	ug/L 100
42) Toluene	7.41	91	108100	10.016	ug/L 100
52) Ethylbenzene	9.04	91	126952	10.620	ug/L 99
53) m,p-Xylene	9.16	106	6989	1.570	ug/L 94
54) o-xylene	9.56	106	11068	2.521	ug/L 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_V\DATA\VV061120\
Data File : VV016856.D
Acq On : 12 Jun 2020 03:09
Operator : SY/MD
Sample : L2915-01DL 250X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F9E04DL

Quant Time: Jun 12 07:46:59 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jun 12 05:55:11 2020
Response via : Initial Calibration

