

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV061120\
 Data File : VV016864.D
 Acq On : 12 Jun 2020 06:43
 Operator : SY/MD
 Sample : VV0611WBL03
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 56 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK82

Quant Time: Jun 12 08:15:47 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jun 12 05:55:11 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	362504	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	339796	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	157058	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	93386	36.51	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	73.02%
7) Chloroethane-d5	1.56	69	55382	36.08	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	72.16%
11) 1,1-Dichloroethene-d2	2.12	63	132084	30.75	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	61.50%
21) 2-Butanone-d5	3.90	46	170401	90.01	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	90.01%
24) Chloroform-d	4.37	84	218566	42.02	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.04%
26) 1,2-Dichloroethane-d4	5.05	65	158849	44.90	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.80%
32) Benzene-d6	5.07	84	419449	42.82	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.64%
36) 1,2-Dichloropropane-d6	6.09	67	131970	43.22	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	86.44%
41) Toluene-d8	7.34	98	384439	42.77	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	85.54%
43) trans-1,3-Dichloropropene-	7.64	79	59358	37.33	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	74.66%
47) 2-Hexanone-d5	8.11	63	95912	72.05	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	72.05%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	163794	40.52	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	81.04%
64) 1,2-Dichlorobenzene-d4	11.65	152	154227	46.01	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.02%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_V\DATA\VV061120\
Data File : VV016864.D
Acq On : 12 Jun 2020 06:43
Operator : SY/MD
Sample : VV0611WBL03
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 56 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK82

Quant Time: Jun 12 08:15:47 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jun 12 05:55:11 2020
Response via : Initial Calibration

