

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV061120\
 Data File : VV016867.D
 Acq On : 12 Jun 2020 09:50
 Operator : SY/MD
 Sample : L2915-21
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 59 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Jun 12 12:27:38 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jun 12 05:55:11 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	340074	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	321743	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	159337	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	107393	44.75	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	89.50%
7) Chloroethane-d5	1.56	69	74625	51.83	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	103.66%
11) 1,1-Dichloroethene-d2	2.12	63	147441	36.59	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.18%
21) 2-Butanone-d5	3.89	46	166626	93.82	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	93.82%
24) Chloroform-d	4.37	84	227610	46.65	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.30%
26) 1,2-Dichloroethane-d4	5.05	65	164435	49.54	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.08%
32) Benzene-d6	5.07	84	441127	47.56	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.12%
36) 1,2-Dichloropropane-d6	6.09	67	137662	47.62	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.24%
41) Toluene-d8	7.34	98	404595	47.54	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.08%
43) trans-1,3-Dichloropropene-	7.64	79	70097	46.55	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	93.10%
47) 2-Hexanone-d5	8.11	63	98066	77.80	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	77.80%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	174010	45.47	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	90.94%
64) 1,2-Dichlorobenzene-d4	11.65	152	175477	51.61	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.22%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_V\DATA\VV061120\
Data File : VV016867.D
Acq On : 12 Jun 2020 09:50
Operator : SY/MD
Sample : L2915-21
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 59 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Jun 12 12:27:38 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jun 12 05:55:11 2020
Response via : Initial Calibration

