

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW061122\
 Data File : VW026123.D
 Acq On : 11 Jun 2022 12:33
 Operator : SY/MD
 Sample : N3247-04DL 5X
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EXYW2DL

Quant Time: Jun 13 01:17:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR060822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Jun 13 01:15:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.622	114	182205	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	177054	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	72802	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.320	65	47124	4.406	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	88.200%	
7) Chloroethane-d5	1.580	69	56469	5.050	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	101.000%	
11) 1,1-Dichloroethene-d2	2.121	63	80090	3.242	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	64.800%	
20) 2-Butanone-d5	3.918	46	89561	56.615	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	113.240%	
24) Chloroform-d	4.362	84	113463	5.005	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	100.200%	
26) 1,2-Dichloroethane-d4	5.043	65	49955	5.403	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	108.000%	
32) Benzene-d6	5.059	84	213140	4.982	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	99.600%	
36) 1,2-Dichloropropane-d6	6.075	67	69535	5.253	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	105.000%	
41) Toluene-d8	7.320	98	160612	4.248	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	85.000%	
43) trans-1,3-Dichloroprop...	7.628	79	14836	3.807	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	76.200%	
46) 2-Hexanone-d5	8.091	63	72868	42.491	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	84.980%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	43409	4.864	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	97.200%	
66) 1,2-Dichlorobenzene-d4	11.622	152	55260	4.961	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	99.200%	
Target Compounds						
					Qvalue	
5) Vinyl chloride	1.323	62	3155	0.150	ug/L	# 62
8) Chloroethane	1.596	64	82485	6.396	ug/L	98
16) Methylene chloride	2.519	84	3355	0.230	ug/L	92
18) trans-1,2-Dichloroethene	2.780	96	1658	0.120	ug/L	96
19) 1,1-Dichloroethane	3.204	63	157115	6.250	ug/L	98
22) cis-1,2-Dichloroethene	3.928	96	7840	0.576	ug/L	88
29) 1,1,1-Trichloroethane	4.619	97	51141	2.201	ug/L	100
34) Trichloroethene	5.921	95	25256	1.737	ug/L	99
47) Tetrachloroethene	7.979	164	1855	0.165	ug/L	95
60) Isopropylbenzene	9.934	105	8953	0.185	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061122\
Data File : VV026123.D
Acq On : 11 Jun 2022 12:33
Operator : SY/MD
Sample : N3247-04DL 5X
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EXYW2DL

Quant Time: Jun 13 01:17:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR060822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Jun 13 01:15:38 2022
Response via : Initial Calibration

