

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW061122\
 Data File : VW026124.D
 Acq On : 11 Jun 2022 12:57
 Operator : SY/MD
 Sample : N3273-05DL 40X
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0B75DL

Quant Time: Jun 13 01:18:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR060822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Jun 13 01:15:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.625	114	171963	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	166160	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	68375	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.320	65	51751	5.127	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	102.600%	
7) Chloroethane-d5	1.584	69	59333	5.622	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	112.400%	
11) 1,1-Dichloroethene-d2	2.121	63	83966	3.601	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	72.000%	
20) 2-Butanone-d5	3.921	46	93075	62.341	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	124.680%	
24) Chloroform-d	4.362	84	117098	5.473	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	109.400%	
26) 1,2-Dichloroethane-d4	5.043	65	49271	5.646	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	113.000%	
32) Benzene-d6	5.060	84	220305	5.487	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	109.800%	
36) 1,2-Dichloropropane-d6	6.076	67	71847	5.784	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	115.600%	
41) Toluene-d8	7.320	98	166533	4.693	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	93.800%	
43) trans-1,3-Dichloroprop...	7.629	79	13521	3.697	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	74.000%	
46) 2-Hexanone-d5	8.092	63	73228	45.501	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	91.000%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	42934	5.126	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	102.600%	
66) 1,2-Dichlorobenzene-d4	11.622	152	55658	5.321	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	106.400%	
Target Compounds						
5) Vinyl chloride	1.327	62	148122	7.485	ug/L	98
16) Methylene chloride	2.523	84	3281	0.238	ug/L	98
18) trans-1,2-Dichloroethene	2.783	96	570	0.044	ug/L	92
22) cis-1,2-Dichloroethene	3.931	96	3192	0.248	ug/L	76
34) Trichloroethene	5.925	95	25416	1.863	ug/L	98
47) Tetrachloroethene	7.982	164	3228	0.306	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061122\
Data File : VV026124.D
Acq On : 11 Jun 2022 12:57
Operator : SY/MD
Sample : N3273-05DL 40X
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0B75DL

Quant Time: Jun 13 01:18:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR060822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Jun 13 01:15:38 2022
Response via : Initial Calibration

