

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW061122\
 Data File : VW026127.D
 Acq On : 11 Jun 2022 14:09
 Operator : SY/MD
 Sample : N3272-14DL 20X
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0B58DL

Quant Time: Jun 13 01:18:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR060822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Jun 13 01:15:38 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.625	114	139344	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	128402	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	50148	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.323	65	45248	5.532	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 110.600%		
7) Chloroethane-d5	1.584	69	55985	6.547	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 131.000%#		
11) 1,1-Dichloroethene-d2	2.124	63	71192	3.768	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 75.400%		
20) 2-Butanone-d5	3.931	46	93655	77.414	ug/L	0.02
Spiked Amount	50.000	Range 40 - 130	Recovery	= 154.820%#		
24) Chloroform-d	4.365	84	112509	6.490	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 129.800%#		
26) 1,2-Dichloroethane-d4	5.043	65	48514	6.861	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 137.200%#		
32) Benzene-d6	5.060	84	200828	6.473	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 129.400%#		
36) 1,2-Dichloropropane-d6	6.076	67	67669	7.050	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 141.000%#		
41) Toluene-d8	7.320	98	143323	5.227	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 104.600%		
43) trans-1,3-Dichloroprop...	7.629	79	13857	4.903	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 98.000%		
46) 2-Hexanone-d5	8.092	63	74190	59.654	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 119.300%		
56) 1,1,2,2-Tetrachloroeth...	10.217	84	41428	6.401	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 128.000%#		
66) 1,2-Dichlorobenzene-d4	11.622	152	48912	6.375	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 127.600%#		
Target Compounds						
						Qvalue
9) Trichlorofluoromethane	1.767	101	12690	0.621	ug/L	100
14) Carbon disulfide	2.310	76	10018	0.313	ug/L	95
16) Methylene chloride	2.523	84	3733	0.334	ug/L	98
34) Trichloroethene	5.924	95	37117	3.521	ug/L	99
47) Tetrachloroethene	7.986	164	877	0.107	ug/L	89
53) m,p-Xylene	9.149	106	869	0.051	ug/L	71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061122\
Data File : VV026127.D
Acq On : 11 Jun 2022 14:09
Operator : SY/MD
Sample : N3272-14DL 20X
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0B58DL

Quant Time: Jun 13 01:18:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR060822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Jun 13 01:15:38 2022
Response via : Initial Calibration

