

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VW061122\  
 Data File : VW026135.D  
 Acq On : 11 Jun 2022 17:19  
 Operator : SY/MD  
 Sample : N3312-04  
 Misc : 25mL/MSVOA\_V/WATER  
 ALS Vial : 21 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 CODY1

Quant Time: Jun 13 01:20:58 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR060822WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Mon Jun 13 01:15:38 2022  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon  | Response | Conc     | Units | Dev(Min) |
|-------------------------------|--------|-------|----------|----------|-------|----------|
| -----                         |        |       |          |          |       |          |
| Internal Standards            |        |       |          |          |       |          |
| 1) 1,4-Difluorobenzene        | 5.622  | 114   | 177591   | 5.000    | ug/L  | 0.00     |
| 28) Chlorobenzene-d5          | 8.853  | 117   | 173251   | 5.000    | ug/L  | 0.00     |
| 58) 1,4-Dichlorobenzene-d4    | 11.249 | 152   | 69721    | 5.000    | ug/L  | 0.00     |
| System Monitoring Compounds   |        |       |          |          |       |          |
| 4) Vinyl Chloride-d3          | 1.320  | 65    | 46978    | 4.506    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 40 - 130 | Recovery | =     | 90.200%  |
| 7) Chloroethane-d5            | 1.580  | 69    | 56681    | 5.201    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 65 - 130 | Recovery | =     | 104.000% |
| 11) 1,1-Dichloroethene-d2     | 2.121  | 63    | 77835    | 3.232    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 60 - 125 | Recovery | =     | 64.600%  |
| 20) 2-Butanone-d5             | 3.915  | 46    | 73324    | 47.555   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000 | Range | 40 - 130 | Recovery | =     | 95.120%  |
| 24) Chloroform-d              | 4.362  | 84    | 113222   | 5.124    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 125 | Recovery | =     | 102.400% |
| 26) 1,2-Dichloroethane-d4     | 5.043  | 65    | 47847    | 5.309    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 130 | Recovery | =     | 106.200% |
| 32) Benzene-d6                | 5.059  | 84    | 212002   | 5.064    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 125 | Recovery | =     | 101.200% |
| 36) 1,2-Dichloropropane-d6    | 6.075  | 67    | 69793    | 5.389    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 60 - 140 | Recovery | =     | 107.800% |
| 41) Toluene-d8                | 7.320  | 98    | 163189   | 4.411    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 130 | Recovery | =     | 88.200%  |
| 43) trans-1,3-Dichloroprop... | 7.628  | 79    | 14943    | 3.919    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 55 - 130 | Recovery | =     | 78.400%  |
| 46) 2-Hexanone-d5             | 8.091  | 63    | 73739    | 43.943   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000 | Range | 45 - 130 | Recovery | =     | 87.880%  |
| 56) 1,1,2,2-Tetrachloroeth... | 10.214 | 84    | 42943    | 4.917    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 65 - 120 | Recovery | =     | 98.400%  |
| 66) 1,2-Dichlorobenzene-d4    | 11.622 | 152   | 55995    | 5.249    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 80 - 120 | Recovery | =     | 105.000% |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VV061122\  
Data File : VV026135.D  
Acq On : 11 Jun 2022 17:19  
Operator : SY/MD  
Sample : N3312-04  
Misc : 25mL/MSVOA\_V/WATER  
ALS Vial : 21 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
CODY1

Quant Time: Jun 13 01:20:58 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR060822WMA.M  
Quant Title : TRACE VOA SFAM1.0  
QLast Update : Mon Jun 13 01:15:38 2022  
Response via : Initial Calibration

