

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW061122\
 Data File : VW026144.D
 Acq On : 11 Jun 2022 21:51
 Operator : SY/MD
 Sample : N3312-11
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0DZ4

Quant Time: Jun 13 07:42:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR060822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Jun 13 07:40:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.625	114	191104	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	182561	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	74428	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.320	65	46790	4.171	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	83.400%
7) Chloroethane-d5	1.581	69	55224	4.709	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.200%
11) 1,1-Dichloroethene-d2	2.121	63	77098	2.975	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	59.600%#
20) 2-Butanone-d5	3.925	46	97350	58.673	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	117.340%
24) Chloroform-d	4.359	84	115298	4.849	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.000%
26) 1,2-Dichloroethane-d4	5.043	65	48613	5.013	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.200%
32) Benzene-d6	5.060	84	213749	4.845	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.000%
36) 1,2-Dichloropropane-d6	6.076	67	68656	5.031	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.600%
41) Toluene-d8	7.320	98	164615	4.223	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.400%
43) trans-1,3-Dichloroprop...	7.629	79	15199	3.783	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	75.600%
46) 2-Hexanone-d5	8.092	63	74509	42.138	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	84.280%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	43217	4.696	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.000%
66) 1,2-Dichlorobenzene-d4	11.622	152	55807	4.901	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061122\
Data File : VV026144.D
Acq On : 11 Jun 2022 21:51
Operator : SY/MD
Sample : N3312-11
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0DZ4

Quant Time: Jun 13 07:42:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR060822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Jun 13 07:40:59 2022
Response via : Initial Calibration

