

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061122\
 Data File : VV026142.D
 Acq On : 11 Jun 2022 21:03
 Operator : SY/MD
 Sample : VV0611WBL02
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK368

Quant Time: Jun 13 07:41:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR060822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Jun 13 07:40:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.622	114	187540	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	183872	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	74670	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.317	65	49253	4.47	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	89.40%
7) Chloroethane-d5	1.577	69	57060	4.96	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.20%
11) 1,1-Dichloroethene-d2	2.117	63	81648	3.21	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	64.20%
20) 2-Butanone-d5	3.924	46	87319	53.63	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	107.26%
24) Chloroform-d	4.358	84	113931	4.88	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.60%
26) 1,2-Dichloroethane-d4	5.040	65	50651	5.32	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.40%
32) Benzene-d6	5.056	84	219673	4.94	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.80%
36) 1,2-Dichloropropane-d6	6.075	67	71627	5.21	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.20%
41) Toluene-d8	7.317	98	168252	4.29	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.80%
43) trans-1,3-Dichloroprop...	7.628	79	14053	3.47	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	69.40%
46) 2-Hexanone-d5	8.091	63	78043	43.82	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	87.64%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	43482	4.69	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.80%
66) 1,2-Dichlorobenzene-d4	11.625	152	58522	5.12	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.40%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061122\
Data File : VV026142.D
Acq On : 11 Jun 2022 21:03
Operator : SY/MD
Sample : VV0611WBL02
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK368

Quant Time: Jun 13 07:41:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR060822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Jun 13 07:40:59 2022
Response via : Initial Calibration

