

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW061122\
 Data File : VW026152.D
 Acq On : 12 Jun 2022 01:01
 Operator : SY/MD
 Sample : N3289-02
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0DR8

Quant Time: Jun 13 07:44:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR060822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Jun 13 07:40:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.625	114	186830	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	180591	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	73857	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.320	65	45355	4.14	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	82.80%
7) Chloroethane-d5	1.581	69	57322	5.00	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	100.00%
11) 1,1-Dichloroethene-d2	2.121	63	76020	3.00	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	60.00%
20) 2-Butanone-d5	3.915	46	69876	43.08	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	86.16%
24) Chloroform-d	4.362	84	114310	4.92	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.40%
26) 1,2-Dichloroethane-d4	5.043	65	48385	5.10	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.00%
32) Benzene-d6	5.060	84	210340	4.82	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.40%
36) 1,2-Dichloropropane-d6	6.076	67	69970	5.18	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.60%
41) Toluene-d8	7.320	98	158170	4.10	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.00%
43) trans-1,3-Dichloroprop...	7.629	79	15150	3.81	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	76.20%
46) 2-Hexanone-d5	8.092	63	76353	43.65	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	87.30%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	44376	4.87	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.40%
66) 1,2-Dichlorobenzene-d4	11.622	152	55278	4.89	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.80%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061122\
Data File : VV026152.D
Acq On : 12 Jun 2022 01:01
Operator : SY/MD
Sample : N3289-02
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0DR8

Quant Time: Jun 13 07:44:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR060822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Jun 13 07:40:59 2022
Response via : Initial Calibration

