

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW061122\
 Data File : VW026153.D
 Acq On : 12 Jun 2022 01:25
 Operator : SY/MD
 Sample : N3289-03
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0DS1

Quant Time: Jun 13 07:44:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR060822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Jun 13 07:40:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.622	114	198796	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	189462	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	76262	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.320	65	46319	3.97	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	79.40%
7) Chloroethane-d5	1.581	69	57516	4.71	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.20%
11) 1,1-Dichloroethene-d2	2.121	63	80736	3.00	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	60.00%
20) 2-Butanone-d5	3.915	46	66448	38.50	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	77.00%
24) Chloroform-d	4.359	84	114172	4.62	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.40%
26) 1,2-Dichloroethane-d4	5.043	65	47787	4.74	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.80%
32) Benzene-d6	5.060	84	217617	4.75	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.00%
36) 1,2-Dichloropropane-d6	6.076	67	69867	4.93	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.60%
41) Toluene-d8	7.320	98	169265	4.18	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.60%
43) trans-1,3-Dichloroprop...	7.632	79	14265	3.42	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	68.40%
46) 2-Hexanone-d5	8.092	63	64041	34.90	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	69.80%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	40380	4.23	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	84.60%
66) 1,2-Dichlorobenzene-d4	11.622	152	55439	4.75	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.00%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061122\
Data File : VV026153.D
Acq On : 12 Jun 2022 01:25
Operator : SY/MD
Sample : N3289-03
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0DS1

Quant Time: Jun 13 07:44:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR060822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Jun 13 07:40:59 2022
Response via : Initial Calibration

