

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW061122\
 Data File : VW026163.D
 Acq On : 12 Jun 2022 05:23
 Operator : SY/MD
 Sample : N3247-11
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Jun 13 07:47:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR060822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Jun 13 07:40:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.625	114	183955	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	180462	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	71738	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.320	65	44298	4.102	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	82.000%
7) Chloroethane-d5	1.584	69	54050	4.788	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	95.800%
11) 1,1-Dichloroethene-d2	2.124	63	75195	3.015	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	60.200%
20) 2-Butanone-d5	3.915	46	66281	41.500	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	83.000%
24) Chloroform-d	4.362	84	111188	4.858	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.200%
26) 1,2-Dichloroethane-d4	5.040	65	46890	5.023	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.400%
32) Benzene-d6	5.059	84	207066	4.748	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.000%
36) 1,2-Dichloropropane-d6	6.075	67	68813	5.101	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.000%
41) Toluene-d8	7.320	98	158699	4.118	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.400%
43) trans-1,3-Dichloroprop...	7.628	79	14034	3.533	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	70.600%
46) 2-Hexanone-d5	8.091	63	66083	37.807	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	75.620%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	40040	4.402	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	88.000%
66) 1,2-Dichlorobenzene-d4	11.622	152	54756	4.989	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061122\
Data File : VV026163.D
Acq On : 12 Jun 2022 05:23
Operator : SY/MD
Sample : N3247-11
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Jun 13 07:47:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR060822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Jun 13 07:40:59 2022
Response via : Initial Calibration

