

Data Path : Z:\voasrv\HPCHEM1\MSVOA V\Data\VV061218\
 Data File : VV006249.D
 Acq On : 12 Jun 2018 16:24
 Operator : SY/MD
 Sample : VSTDCCC005EC
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 LabSampleId :
 VSTD00562

Quant Time: Jun 13 01:29:58 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR053118WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Jun 13 01:21:06 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Difluorobenzene	5.000	5.000	0.0	98	0.00
2 T	Dichlorodifluoromethane	5.000	4.939	1.2	95	0.00
3 T	Chloromethane	5.000	5.412	-8.2	106	0.00
4 S	Vinyl Chloride-d3	5.000	4.926	1.5	96	0.00
5 T	Vinyl chloride	5.000	5.257	-5.1	101	0.00
6 T	Bromomethane	5.000	4.971	0.6	98	0.00
7 S	Chloroethane-d5	5.000	5.665	-13.3	112	0.00
8 T	Chloroethane	5.000	5.829	-16.6	111	0.00
9 T	Trichlorofluoromethane	5.000	5.773	-15.5	109	0.00
10 T	1,1,2-Trichloro-1,2,2-trifl	5.000	5.824	-16.5	110	0.00
11 S	1,1-Dichloroethene-d2	5.000	5.099	-2.0	98	0.00
12 T	1,1-Dichloroethene	5.000	5.808	-16.2	110	0.00
13 T	Acetone	50.000	72.448	-44.9#	153	0.00
14 T	Carbon disulfide	5.000	5.099	-2.0	97	0.00
15 T	Methyl Acetate	5.000	6.576	-31.5#	142	0.00
16 T	Methylene chloride	5.000	5.750	-15.0	118	0.00
17 T	Methyl tert-butyl Ether	5.000	5.781	-15.6	109	0.00
18 T	trans-1,2-Dichloroethene	5.000	5.854	-17.1	110	0.00
19 T	1,1-Dichloroethane	5.000	5.642	-12.8	108	0.00
20 S	2-Butanone-d5	50.000	62.646	-25.3	135	0.00
21 T	2-Butanone	50.000	61.194	-22.4	123	0.00
22 T	cis-1,2-Dichloroethene	5.000	5.586	-11.7	106	0.00
23 T	Bromochloromethane	5.000	5.559	-11.2	105	0.00
24 S	Chloroform-d	5.000	5.312	-6.2	103	0.00
25 T	Chloroform	5.000	5.635	-12.7	108	0.00
26 S	1,2-Dichloroethane-d4	5.000	5.240	-4.8	103	0.00
27 T	1,2-Dichloroethane	5.000	5.568	-11.4	106	0.00
28 I	Chlorobenzene-d5	5.000	5.000	0.0	99	0.00
29 T	1,1,1-Trichloroethane	5.000	5.412	-8.2	102	0.00
30 T	Cyclohexane	5.000	5.343	-6.9	102	0.00
31 T	Carbon tetrachloride	5.000	5.337	-6.7	102	0.00
32 S	Benzene-d6	5.000	5.030	-0.6	99	0.00
33 T	Benzene	5.000	5.497	-9.9	105	0.00
34 T	Trichloroethene	5.000	5.485	-9.7	104	0.00
35 T	Methylcyclohexane	5.000	5.371	-7.4	102	0.00
36 S	1,2-Dichloropropane-d6	5.000	5.198	-4.0	103	0.00
37 T	1,2-Dichloropropane	5.000	5.474	-9.5	105	0.00
38 T	Bromodichloromethane	5.000	5.259	-5.2	99	0.00
39 T	cis-1,3-Dichloropropene	5.000	5.265	-5.3	100	0.00
40 T	4-Methyl-2-pentanone	50.000	55.261	-10.5	105	0.00
41 S	Toluene-d8	5.000	4.833	3.3	97	0.00
42 T	Toluene	5.000	5.485	-9.7	105	0.00
43 S	trans-1,3-Dichloropropene-d	5.000	4.750	5.0	96	0.00
44 T	trans-1,3-Dichloropropene	5.000	5.200	-4.0	100	0.00
45 T	1,1,2-Trichloroethane	5.000	5.508	-10.2	106	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA V\Data\VV061218\
 Data File : VV006249.D
 Acq On : 12 Jun 2018 16:24
 Operator : SY/MD
 Sample : VSTDCCC005EC
 Misc : 25 mL/MSVOA V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 LabSampleId :
 VSTD00562

Quant Time: Jun 13 01:29:58 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR053118WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Jun 13 01:21:06 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
46 S	2-Hexanone-d5	50.000	53.825	-7.7	104	0.00
47 T	Tetrachloroethene	5.000	5.718	-14.4	108	0.00
48 T	2-Hexanone	50.000	56.276	-12.6	105	0.00
49 T	Dibromochloromethane	5.000	5.223	-4.5	101	0.00
50 T	1,2-Dibromoethane	5.000	5.517	-10.3	105	0.00
51 T	Chlorobenzene	5.000	5.623	-12.5	107	0.00
52 T	Ethylbenzene	5.000	5.497	-9.9	104	0.00
53 T	m,p-xylene	5.000	5.490	-9.8	106	0.00
54 T	o-xylene	5.000	5.550	-11.0	105	0.00
55 T	Styrene	5.000	5.493	-9.9	104	0.00
56 T	Isopropylbenzene	5.000	5.585	-11.7	106	0.00
57 S	1,1,2,2-Tetrachloroethane-d	5.000	5.334	-6.7	105	0.00
58 T	1,1,2,2-Tetrachloroethane	5.000	5.123	-2.5	99	0.00
59	1,2,3-Trichloropropane	5.000	5.598	-12.0	106	0.00
60 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	103	0.00
61 T	Bromoform	5.000	4.815	3.7	95	0.00
62 T	1,3-Dichlorobenzene	5.000	5.474	-9.5	109	0.00
63 T	1,4-Dichlorobenzene	5.000	5.492	-9.8	110	0.00
64 S	1,2-Dichlorobenzene-d4	5.000	5.260	-5.2	108	0.00
65 T	1,2-Dichlorobenzene	5.000	5.571	-11.4	110	0.00
66 T	1,2-Dibromo-3-chloropropane	5.000	4.804	3.9	93	0.00
67	1,3,5-Trichlorobenzene	5.000	5.621	-12.4	110	0.00
68 T	1,2,4-trichlorobenzene	5.000	5.632	-12.6	108	0.00
69	Naphthalene	5.000	5.314	-6.3	99	0.00
70 T	1,2,3-Trichlorobenzene	5.000	5.668	-13.4	109	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0