

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061220\
 Data File : VV016875.D
 Acq On : 12 Jun 2020 13:38
 Operator : SY/MD
 Sample : L2848-20
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AH5

Quant Time: Jun 13 02:02:01 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Jun 13 01:34:58 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	249998	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	235385	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	115825	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	79837	45.26	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	90.52%
7) Chloroethane-d5	1.55	69	52197	49.31	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	98.62%
11) 1,1-Dichloroethene-d2	2.11	63	104955	35.43	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	70.86%
21) 2-Butanone-d5	3.89	46	144529	110.70	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	110.70%
24) Chloroform-d	4.37	84	171033	47.68	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.36%
26) 1,2-Dichloroethane-d4	5.05	65	128679	52.74	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.48%
32) Benzene-d6	5.07	84	333592	49.16	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.32%
36) 1,2-Dichloropropane-d6	6.09	67	104812	49.55	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.10%
41) Toluene-d8	7.34	98	297479	47.78	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.56%
43) trans-1,3-Dichloropropene-	7.64	79	51638	46.88	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	93.76%
47) 2-Hexanone-d5	8.11	63	74888	81.21	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	81.21%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	130780	46.71	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	93.42%
64) 1,2-Dichlorobenzene-d4	11.65	152	129272	52.30	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	104.60%

Target Compounds					Ovalue
13) Acetone	2.17	43	5612	6.524 ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV061220\
Data File : VV016875.D
Acq On : 12 Jun 2020 13:38
Operator : SY/MD
Sample : L2848-20
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AH5

Quant Time: Jun 13 02:02:01 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Jun 13 01:34:58 2020
Response via : Initial Calibration

