

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061220\
 Data File : VV016877.D
 Acq On : 12 Jun 2020 14:26
 Operator : SY/MD
 Sample : L2848-06
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK02

Quant Time: Jun 13 02:09:39 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Jun 13 01:34:58 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	94236	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	92840	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	42707	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	67086	100.89	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	201.78%#
7) Chloroethane-d5	1.56	69	47723	119.61	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	239.22%#
11) 1,1-Dichloroethene-d2	2.12	63	93261	83.52	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	167.04%#
21) 2-Butanone-d5	3.90	46	109972	223.45	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	223.45%#
24) Chloroform-d	4.37	84	147880	109.37	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	218.74%#
26) 1,2-Dichloroethane-d4	5.05	65	105737	114.97	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	229.94%#
32) Benzene-d6	5.07	84	285435	106.66	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	213.32%#
36) 1,2-Dichloropropane-d6	6.09	67	87411	104.78	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	209.56%#
41) Toluene-d8	7.34	98	255268	103.94	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	207.88%#
43) trans-1,3-Dichloropropene-	7.64	79	43084	99.16	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	198.32%#
47) 2-Hexanone-d5	8.11	63	55524	152.67	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	152.67%#
57) 1,1,2,2-Tetrachloroethane-	10.24	84	105804	95.81	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	191.62%#
64) 1,2-Dichlorobenzene-d4	11.65	152	107990	118.49	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	236.98%#

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV061220\
Data File : VV016877.D
Acq On : 12 Jun 2020 14:26
Operator : SY/MD
Sample : L2848-06
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK02

Quant Time: Jun 13 02:09:39 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Jun 13 01:34:58 2020
Response via : Initial Calibration

