

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061220\
 Data File : VV016899.D
 Acq On : 12 Jun 2020 23:20
 Operator : SY/MD
 Sample : L2915-08 40X
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F9E16

Quant Time: Jun 13 05:02:54 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Jun 13 03:02:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	372626	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	350781	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	169214	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	100000	38.03	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	76.06%
7) Chloroethane-d5	1.57	69	57614	36.52	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	73.04%
11) 1,1-Dichloroethene-d2	2.12	63	145276	32.90	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	65.80%
21) 2-Butanone-d5	3.90	46	175452	90.16	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	90.16%
24) Chloroform-d	4.37	84	227999	42.64	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.28%
26) 1,2-Dichloroethane-d4	5.05	65	167915	46.17	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.34%
32) Benzene-d6	5.07	84	444942	44.00	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.00%
36) 1,2-Dichloropropane-d6	6.09	67	135194	42.89	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	85.78%
41) Toluene-d8	7.34	98	399224	43.02	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	86.04%
43) trans-1,3-Dichloropropene-	7.64	79	67429	41.07	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	82.14%
47) 2-Hexanone-d5	8.11	63	103216	75.11	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	75.11%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	172870	41.43	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	82.86%
64) 1,2-Dichlorobenzene-d4	11.65	152	162643	45.04	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.08%

Target Compounds

					Ovalue
29) Cyclohexane	4.71	56	9377	2.151 ug/L	95
33) Benzene	5.12	78	2458158	236.849 ug/L	100
42) Toluene	7.41	91	305338	27.908 ug/L	100
52) Ethylbenzene	9.04	91	186889	15.422 ug/L	97
53) m,p-Xylene	9.16	106	16586	3.676 ug/L	97
54) o-xylene	9.57	106	32385	7.277 ug/L	93
56) Isopropylbenzene	9.95	105	7201	0.602 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV061220\
Data File : VV016899.D
Acq On : 12 Jun 2020 23:20
Operator : SY/MD
Sample : L2915-08 40X
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F9E16

Quant Time: Jun 13 05:02:54 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Jun 13 03:02:20 2020
Response via : Initial Calibration

