

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061220\
 Data File : VV016905.D
 Acq On : 13 Jun 2020 02:30
 Operator : SY/MD
 Sample : L2915-21
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Jun 13 06:32:32 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Jun 13 06:09:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	338886	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	323489	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	159910	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	84775	35.45	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	70.90%
7) Chloroethane-d5	1.56	69	64041	44.63	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	89.26%
11) 1,1-Dichloroethene-d2	2.11	63	125386	31.23	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	62.46%
21) 2-Butanone-d5	3.89	46	165679	93.61	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	93.61%
24) Chloroform-d	4.37	84	214655	44.15	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.30%
26) 1,2-Dichloroethane-d4	5.05	65	159560	48.24	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.48%
32) Benzene-d6	5.07	84	410602	44.03	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.06%
36) 1,2-Dichloropropane-d6	6.09	67	127787	43.96	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	87.92%
41) Toluene-d8	7.34	98	371839	43.45	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	86.90%
43) trans-1,3-Dichloropropene-	7.64	79	63964	42.25	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	84.50%
47) 2-Hexanone-d5	8.11	63	99915	78.84	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	78.84%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	170682	44.36	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	88.72%
64) 1,2-Dichlorobenzene-d4	11.65	152	165974	48.64	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.28%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV061220\
Data File : VV016905.D
Acq On : 13 Jun 2020 02:30
Operator : SY/MD
Sample : L2915-21
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Jun 13 06:32:32 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Jun 13 06:09:05 2020
Response via : Initial Calibration

