

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV061220\
 Data File : VV016883.D
 Acq On : 12 Jun 2020 16:59
 Operator : SY/MD
 Sample : VV0612WBL02
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK84

Quant Time: Jun 13 03:17:31 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Jun 13 03:02:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	345782	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	329082	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	160449	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	103673	42.49	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	84.98%
7) Chloroethane-d5	1.58	69	78070	53.33	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	106.66%
11) 1,1-Dichloroethene-d2	2.12	63	150497	36.73	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.46%
21) 2-Butanone-d5	3.89	46	179159	99.21	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	99.21%
24) Chloroform-d	4.37	84	237368	47.84	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.68%
26) 1,2-Dichloroethane-d4	5.05	65	172843	51.22	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.44%
32) Benzene-d6	5.07	84	454545	47.92	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.84%
36) 1,2-Dichloropropane-d6	6.09	67	139903	47.31	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	94.62%
41) Toluene-d8	7.34	98	415010	47.67	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.34%
43) trans-1,3-Dichloropropene-	7.64	79	70320	45.66	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	91.32%
47) 2-Hexanone-d5	8.11	63	104347	80.94	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	80.94%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	179063	45.74	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.48%
64) 1,2-Dichlorobenzene-d4	11.65	152	171471	50.08	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.16%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV061220\
Data File : VV016883.D
Acq On : 12 Jun 2020 16:59
Operator : SY/MD
Sample : VV0612WBL02
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK84

Quant Time: Jun 13 03:17:31 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM060820WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Jun 13 03:02:20 2020
Response via : Initial Calibration

