

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW061222\
 Data File : VW026177.D
 Acq On : 12 Jun 2022 21:55
 Operator : SY/MD
 Sample : N3278-01 200X
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0BC8

Quant Time: Jun 13 06:37:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061222WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Jun 13 02:52:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.625	114	177355	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	171513	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	66003	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.320	65	36952	1.593	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	31.800%#	
7) Chloroethane-d5	1.584	69	47359	2.592	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	51.800%#	
11) 1,1-Dichloroethene-d2	2.124	63	61951	1.661	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	33.200%#	
20) 2-Butanone-d5	3.915	46	77376	44.718	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	89.440%	
24) Chloroform-d	4.362	84	103345	3.142	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	62.800%#	
26) 1,2-Dichloroethane-d4	5.043	65	45915	3.462	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	69.200%#	
32) Benzene-d6	5.060	84	187606	2.806	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	56.200%#	
36) 1,2-Dichloropropane-d6	6.076	67	62893	3.262	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	65.200%	
41) Toluene-d8	7.320	98	141201	2.372	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	47.400%#	
43) trans-1,3-Dichloroprop...	7.632	79	14325	2.461	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	49.200%#	
46) 2-Hexanone-d5	8.092	63	70753	39.045	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	78.100%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	39868	3.470	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	69.400%	
66) 1,2-Dichlorobenzene-d4	11.622	152	51729	3.496	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	70.000%#	
Target Compounds						Qvalue
9) Trichlorofluoromethane	1.770	101	4940	0.198	ug/L	97
34) Trichloroethene	5.924	95	28490	2.034	ug/L	96
47) Tetrachloroethene	7.979	164	17846	1.774	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061222\
Data File : VV026177.D
Acq On : 12 Jun 2022 21:55
Operator : SY/MD
Sample : N3278-01 200X
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0BC8

Quant Time: Jun 13 06:37:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061222WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Jun 13 02:52:24 2022
Response via : Initial Calibration

