

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW061222\
 Data File : VW026189.D
 Acq On : 13 Jun 2022 02:41
 Operator : SY/MD
 Sample : N3278-13 200X
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0BF4

Quant Time: Jun 13 06:39:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061222WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Jun 13 02:52:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.625	114	180174	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	175711	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	69137	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.320	65	39981	1.697	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	34.000%#	
7) Chloroethane-d5	1.584	69	50958	2.746	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	55.000%#	
11) 1,1-Dichloroethene-d2	2.124	63	66624	1.759	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	35.200%#	
20) 2-Butanone-d5	3.918	46	71650	40.761	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	81.520%	
24) Chloroform-d	4.362	84	109873	3.288	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	65.800%#	
26) 1,2-Dichloroethane-d4	5.044	65	47660	3.538	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	70.800%	
32) Benzene-d6	5.060	84	196205	2.864	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	57.200%#	
36) 1,2-Dichloropropane-d6	6.076	67	67026	3.393	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	67.800%	
41) Toluene-d8	7.320	98	149039	2.443	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	48.800%#	
43) trans-1,3-Dichloroprop...	7.632	79	14131	2.370	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	47.400%#	
46) 2-Hexanone-d5	8.092	63	79144	42.632	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	85.260%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	42375	3.601	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	72.000%	
66) 1,2-Dichlorobenzene-d4	11.622	152	53640	3.461	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	69.200%#	
Target Compounds						
					Qvalue	
9) Trichlorofluoromethane	1.767	101	110898	4.374	ug/L	98
16) Methylene chloride	2.523	84	2871	0.202	ug/L	93
22) cis-1,2-Dichloroethene	3.931	96	6960	0.524	ug/L	91
31) Carbon tetrachloride	4.841	117	1788	0.096	ug/L	97
34) Trichloroethene	5.921	95	7546179	525.774	ug/L	95
47) Tetrachloroethene	7.976	164	260233	25.249	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061222\
Data File : VV026189.D
Acq On : 13 Jun 2022 02:41
Operator : SY/MD
Sample : N3278-13 200X
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0BF4

Quant Time: Jun 13 06:39:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061222WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Jun 13 02:52:24 2022
Response via : Initial Calibration

