

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW061222\
 Data File : VW026190.D
 Acq On : 13 Jun 2022 03:04
 Operator : SY/MD
 Sample : N3278-14 200X
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0BF5

Quant Time: Jun 13 06:40:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061222WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Jun 13 02:52:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.625	114	180021	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	171234	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	67023	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.323	65	39734	1.688	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	33.800%#	
7) Chloroethane-d5	1.584	69	49031	2.644	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	52.800%#	
11) 1,1-Dichloroethene-d2	2.124	63	66690	1.762	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	35.200%#	
20) 2-Butanone-d5	3.915	46	76587	43.606	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	87.220%	
24) Chloroform-d	4.365	84	108728	3.257	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	65.200%#	
26) 1,2-Dichloroethane-d4	5.043	65	48140	3.576	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	71.600%	
32) Benzene-d6	5.059	84	194526	2.914	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	58.200%#	
36) 1,2-Dichloropropane-d6	6.075	67	66856	3.473	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	69.400%	
41) Toluene-d8	7.320	98	144194	2.426	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	48.600%#	
43) trans-1,3-Dichloroprop...	7.628	79	13660	2.351	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	47.000%#	
46) 2-Hexanone-d5	8.091	63	72161	39.887	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	79.780%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	41083	3.582	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	71.600%	
66) 1,2-Dichlorobenzene-d4	11.625	152	52715	3.509	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	70.200%#	
Target Compounds						
16) Methylene chloride	2.526	84	3310	0.233	ug/L	93
34) Trichloroethene	5.924	95	21402	1.530	ug/L	98
47) Tetrachloroethene	7.985	164	1289	0.128	ug/L	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061222\
Data File : VV026190.D
Acq On : 13 Jun 2022 03:04
Operator : SY/MD
Sample : N3278-14 200X
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0BF5

Quant Time: Jun 13 06:40:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061222WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Jun 13 02:52:24 2022
Response via : Initial Calibration

