

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW061223\
 Data File : VW031501.D
 Acq On : 12 Jun 2023 11:57
 Operator : SY/MD
 Sample : 03101-02ME
 Misc : 11.88g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EZEL7ME

Quant Time: Jun 13 01:42:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM060723WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Jun 13 01:41:31 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.538	114	218422	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.792	117	221576	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	114402	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	56468	29.384	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	58.760%#
7) Chloroethane-d5	1.516	69	30803	19.798	ug/L	-0.02
Spiked Amount	50.000	Range	70 - 130	Recovery	=	39.600%#
11) 1,1-Dichloroethene-d2	2.043	65	35025	38.455	ug/L	-0.02
Spiked Amount	50.000	Range	60 - 125	Recovery	=	76.920%
21) 2-Butanone-d5	3.809	46	128863	92.785	ug/L	0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	92.780%
24) Chloroform-d	4.256	84	162410	44.429	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.860%
26) 1,2-Dichloroethane-d4	4.947	65	117428	47.885	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.760%
32) Benzene-d6	4.963	84	283885	43.396	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.800%
36) 1,2-Dichloropropane-d6	5.995	67	106306	46.751	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	93.500%
41) Toluene-d8	7.249	98	236739	39.395	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	78.780%#
43) trans-1,3-Dichloroprop...	7.561	79	45958	40.124	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	80.240%
47) 2-Hexanone-d5	8.053	63	70641	82.483	ug/L	0.02
Spiked Amount	100.000	Range	45 - 130	Recovery	=	82.480%
56) 1,1,2,2-Tetrachloroeth...	10.165	84	126790	44.183	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	88.360%
66) 1,2-Dichlorobenzene-d4	11.570	152	103107	48.585	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.160%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061223\
Data File : VV031501.D
Acq On : 12 Jun 2023 11:57
Operator : SY/MD
Sample : 03101-02ME
Misc : 11.88g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EZEL7ME

Quant Time: Jun 13 01:42:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM060723WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Jun 13 01:41:31 2023
Response via : Initial Calibration

