

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV061225\
 Data File : VV038790.D
 Acq On : 12 Jun 2025 13:19
 Operator : SY/MD
 Sample : VV0612WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK287

Quant Time: Jun 13 12:41:19 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR060425WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jun 13 10:50:34 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	209073	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.776	117	220337	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.175	152	89867	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.288	65	110295	3.768	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	75.400%	
7) Chloroethane-d5	1.542	69	87033	4.201	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	84.000%	
11) 1,1-Dichloroethene-d2	2.066	65	43126	3.638	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	72.800%	
20) 2-Butanone-d5	3.790	46	179151	47.588	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	95.180%	
24) Chloroform-d	4.249	84	188313	4.519	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	90.400%	
26) 1,2-Dichloroethane-d4	4.941	65	88457	4.659	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	93.200%	
32) Benzene-d6	4.960	84	322227	4.065	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	81.400%	
36) 1,2-Dichloropropane-d6	5.982	67	111468	4.522	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	90.400%	
41) Toluene-d8	7.240	98	235908	3.414	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	68.200%#	
43) trans-1,3-Dichloroprop...	7.555	79	32582	3.814	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	76.200%	
46) 2-Hexanone-d5	8.014	63	131802	45.883	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	91.760%	
56) 1,1,2,2-Tetrachloroeth...	10.143	84	81177	4.530	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	90.600%	
66) 1,2-Dichlorobenzene-d4	11.551	152	78870	4.746	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	95.000%	
Target Compounds						
16) Methylene chloride	2.452	84	5777	0.241	ug/L	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV061225\
Data File : VV038790.D
Acq On : 12 Jun 2025 13:19
Operator : SY/MD
Sample : VV0612WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK287

Quant Time: Jun 13 12:41:19 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR060425WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jun 13 10:50:34 2025
Response via : Initial Calibration

