

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW061322\
 Data File : VW026214.D
 Acq On : 13 Jun 2022 17:10
 Operator : SY/MD
 Sample : N3272-15
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0B59

Quant Time: Jun 14 02:40:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061222WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Jun 14 02:36:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.622	114	162215	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	158665	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	64991	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.314	65	92522	4.361	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	87.200%
7) Chloroethane-d5	1.577	69	74996	4.489	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.800%
11) 1,1-Dichloroethene-d2	2.114	63	120986	3.547	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	71.000%
20) 2-Butanone-d5	3.912	46	98712	62.373	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	124.740%
24) Chloroform-d	4.359	84	142107	4.724	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.400%
26) 1,2-Dichloroethane-d4	5.040	65	59683	4.921	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.400%
32) Benzene-d6	5.056	84	273013	4.414	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.200%
36) 1,2-Dichloropropane-d6	6.072	67	86396	4.844	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.800%
41) Toluene-d8	7.317	98	222184	4.034	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.600%
43) trans-1,3-Dichloroprop...	7.625	79	23800	4.421	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	88.400%
46) 2-Hexanone-d5	8.092	63	89676	53.495	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	107.000%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	52058	4.898	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.000%
66) 1,2-Dichlorobenzene-d4	11.625	152	71897	4.935	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.800%
Target Compounds						
					Qvalue	
9) Trichlorofluoromethane	1.761	101	104950	4.598	ug/L	99
31) Carbon tetrachloride	4.838	117	3455	0.205	ug/L	94
34) Trichloroethene	5.915	95	447608	34.537	ug/L	99
47) Tetrachloroethene	7.976	164	131304	14.108	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV061322\
Data File : VV026214.D
Acq On : 13 Jun 2022 17:10
Operator : SY/MD
Sample : N3272-15
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0B59

Quant Time: Jun 14 02:40:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR061222WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Jun 14 02:36:38 2022
Response via : Initial Calibration

